## Store at -20°C

#29118

# SimpleChIP<sup>®</sup> Human EIF4G1 Intron 8 Primers

Small 500 µl (250 PCR reactions)



### Support: +1-978-867-2388 (U.S.) www.cellsignal.com/support

Orders: 877-616-2355 (U.S.) orders@cellsignal.com

Entrez-Gene ID #1981 UniProt ID #Q04637

## For Research Use Only. Not For Use In Diagnostic Procedures.

Applications	Species Cross-Reactivity		neal/Extension	PCR	Product Leng	Ith
ChIP	Н	(	i5°C		100 bp	
contain a mix of forward an specific to intron 8 of the hi factor 4 γ gene. These prim that has been isolated using (ChIP). Primers have been Universal qPCR Master Mi»	<sup>®</sup> Human EIF4G1 Intron 8 Primers Id reverse PCR primers that are uman eukaryotic translation initiatic ters can be used to amplify DNA g chromatin immunoprecipitations optimized for use with SimpleChIP <sup>0</sup> x #88989 and have been tested in IP <sup>®</sup> Plus Enzymatic Chromatin IP	2 2 2 2 2 2 2 2	8 7 6 5 4	1	0	• • 1 2

SimpleChIP® Human EIF4G1 Intron 8 Primers were tested on DNA isolated from cross-linked cells using the SimpleChIP® Plus Enzymatic Chromatin IP Kit (Magnetic Beads) #9005. Real-time PCR was performed in duplicate on a serial dilution of 2% total input DNA (20 ng, 4 ng, 0.8 ng, and 0.16 ng) using a real-time PCR detection system and SimpleChIP® Universal qPCR Master Mix #88989. The PCR amplification efficiency (E) and correlation coefficient (R<sup>2</sup>) were calculated based on the corresponding threshold cycle (C<sub>1</sub>) of each dilution sample during 40 cycles of real-time PCR (95°C denaturation for 15 sec, 65°C anneal/extension for 60 sec).

**Storage:** Supplied in nuclease-free water at a concentration of 5  $\mu$ M (each primer is at a final concentration of 5  $\mu$ M). *Store at -20°C*.

## **Directions for Use:**

New 10/18

- Label the appropriate number of PCR tubes or PCR plates compatible with the model of real-time PCR machine to be used. PCR reactions should be performed in duplicate and should include a tube with no DNA to control for contamination, and a serial dilution of a 2% total input chromatin DNA (undiluted, 1:5, 1:25, 1:125), which is used to create a standard curve and determine amplification efficiency.
- Add 2 µl of the appropriate ChIP DNA sample to each tube or well of the PCR plate.
- B. Prepare a master PCR reaction mix as described below. Add enough reagents for two extra reactions to account for loss of volume. Add 18 µl of the master PCR reaction mix to each PCR reaction tube or well of the PCR plate.


Reagent	Volume for 1 PCR Reaction (2	20 µl)
Nuclease-free H <sub>2</sub>	0	6 µl
5 µM SimpleChl	P <sup>®</sup> Primers	2 µl
SimpleChIP® Un	iversal qPCR Master Mix #88989	10 µl

4. Start the following PCR reaction program:

- a. Initial Denaturation: 95°C for 3 min.
- b. Denaturation: 95°C for 15 sec.
- c. Anneal and Extension: Primer-specific temp. for 60 sec.
- d. Repeat steps b and c for a total of 40 cycles.
- Analyze quantitative PCR results using software provided with the real-time PCR machine.

Thank you for your recent purchase. If you would like to provide a review visit cellsignal.com/comments.





PCR product melting curves were obtained for real-time PCR reactions performed using SimpleChIP<sup>®</sup> Human EIF4G1 Intron 8 Primers. Data is shown for both duplicate PCR reactions using 20 ng of total DNA. The melt curve consists of 80 melt cycles, starting at 55°C with increments of 0.5°C per cycle. Each peak is formed from the degradation of a single PCR product.

#### @ 2018 Cell Signaling Technology, Inc.

SimpleChIP and Cell Signaling Technology are trademarks of Cell Signaling Technology, Inc.

Applications: W—Western IP—Immunoprecipitation IHC—Immunohistochemistry ChIP—Chromatin Immunoprecipitation IF—Immunofluorescence F—Flow cytometry E-P—ELISA-Peptide Species Cross-Reactivity: H—human M—mouse R—rat Hm—hamster Mk—monkey Mi—mink C—chicken Dm—D, melanogaster X—Xenopus Z—zebrafish B—bovine Dg—dog Pg—pig Sc—S, cerevisiae Ce—C, elegans Hr—Horse AII—all species expected Species enclosed in parentheses are predicted to react based on 100% homology