SimpleChIP® Human FOXN3 **Intron 3 Primers**

500 ul (250 PCR reactions)



Support: +1-978-867-2388 (U.S.) www.cellsignal.com/support

> Orders: 877-616-2355 (U.S.) orders@cellsignal.com


Entrez-Gene ID #1112 **UniProt ID** #000409

New 10/18


For Research Use Only. Not For Use In Diagnostic Procedures.


Applications	Species Cross-Reactivity	Primer Anneal/Extension	PCR Product Length
ChIP	Н	65°C	94 bp

Description: SimpleChIP® Human FOXN3 Intron 3 Primers contain a mix of forward and reverse PCR primers that are specific to intron 3 of the human forkhead box N3 (FOXN3) gene. These primers can be used to amplify DNA that has been isolated using chromatin immunoprecipitation (ChIP). Primers have been optimized for use in quantitative real-time PCR and have been tested in conjunction with SimpleChIP® Enzymatic Chromatin IP Kits #9004 and #9005, ChIP-validated antibodies from Cell Signaling Technology®, and SimpleChlP® Universal aPCR Master Mix #88989.



isolated from cross-linked cells using the SimpleChIP® Plus Enzy-(20 ng, 4 ng, 0.8 ng, and 0.16 ng) using a real-time PCR detection dilution sample during 40 cycles of real-time PCR (95°C denaturation for 15 sec, 65°C anneal/extension for 60 sec).





PCR product melting curves were obtained for real-time PCR reactions performed using SimpleChIP® Human FOXN3 Intron 3 Primers. Data is shown for both duplicate PCR reactions using 20 ng of total DNA. The melt curve consists of 80 melt cycles, starting at 55°C with increments of 0.5°C per cycle. Each peak is formed from the degradation of a single PCR product.

Storage: Supplied in nuclease-free water at a concentration of 5 μM (each primer is at a final concentration of 5 μM). Store at -20°C.

Directions for Use:

- 1. Label the appropriate number of PCR tubes or PCR plates compatible with the model of real-time PCR machine to be used. PCR reactions should be performed in duplicate and should include a tube with no DNA to control for contamination, and a serial dilution of a 2% total input chromatin DNA (undiluted, 1:5, 1:25, 1:125), which is used to create a standard curve and determine amplification efficiency.
- 2. Add 2 µl of the appropriate ChIP DNA sample to each tube or well of the PCR plate.
- 3. Prepare a master PCR reaction mix as described below. Add enough reagents for two extra reactions to account for loss of volume. Add 18 µl of the master PCR reaction mix to each PCR reaction tube or well of the PCR plate.

Reagent Volume for 1 PCR Reaction (20 µl)

Nuclease-free H₂O	6 µl
5 μM SimpleChĺ̄P® Primers	2 μΙ
SimpleChIP® Universal gPCR Master Mix #88989	10 ul

- 4. Start the following PCR reaction program:
 - a. Initial Denaturation: 95°C for 3 min.
 - b. Denaturation: 95°C for 15 sec.
 - c. Anneal and Extension: Primer-specific temp. for 60 sec.
 - d. Repeat steps b and c for a total of 40 cycles.
- 5. Analyze quantitative PCR results using software provided with the real-time PCR machine.

Thank you for your recent purchase. If you would like to provide a review visit cellsignal.com/comments.

www.cellsignal.com